

Validated Blood Pressure Devices

The ultimate judgment regarding whether a BP measurement device meets the requisite VDL Criteria rests with the Independent Review Committee and is not in any way determined or influenced by the AMA. The AMA does not receive funding from any device manufacturer or other third party in relation to the development of the VDL Criteria or VDL process.



iHealth Track Blood Pressure Monitor (KN-550BT)

22cm-42cm

Brand	iHealth
Cuff Type	Adult , Large , Small Adult
Device Type	Home
Connectivity	Bluetooth , Smartphone App
Device Type	Home
Population Served	General Adult
Price	\$
Validation Protocol	ANSI/AAMI/ISO 81060-2: 2018



NOTE: Additional devices may be added to the VDL in the future through successive application and review processes as more devices are submitted and reviewed by the Independent Review Committee. The listing currently contains devices that measure other biometrics, it is important to note that for those devices, only the BP measurement components have been reviewed for validation. Clinical Disclaimer: Devices on this list for use at home are validated for clinical accuracy for the population tested and may not always provide accurate BP measurements for a specific individual. Whenever practical, patients are encouraged to bring their home BP devices to the physician's office or some other qualified health care facility for comparison with an accurate device.

*The American Medical Association ("AMA") does not receive funding from any third-party in relation to the development of the Validated Device Listing Criteria ("VDLTM Criteria"), and does not receive funding from any blood pressure ("BP") device manufacturer or other third-party in relation to the development of the VDL process. No recommendation, promotion, or endorsement is implied or intended by the AMA (or any of the AMA's affiliated or partner organizations) of any third-party organization, product, device, policy, or service. BP measurement devices are selected to appear on the VDL through an independent review process that determines which available BP measurement devices meet established VDL Criteria for the validation of clinical accuracy. An Independent Review Committee, composed of members who are experts in the BP field, assesses whether a BP measurement device satisfies the VDL Criteria. The ultimate judgment regarding whether a BP measurement device meets the requisite VDL Criteria rests with the Independent Review Committee, and is not in any way determined or influenced by the AMA.